

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/26/2015**

HMS
mmB

Instrument Serial Number:	010610	As of 26-Jun-15
----------------------------------	---------------	------------------------

Start Date	End Date	Agency Name
25-Jun-15		Arlington County PD
12-Jun-15	25-Jun-15	DFS Central Lab
14-Jan-15	12-Jun-15	Cumberland County Sheriff's Office
16-Dec-14	14-Jan-15	DFS Central Lab
15-Jul-14	16-Dec-14	Winchester PD
18-Jun-14	15-Jul-14	DFS Central Lab
02-Aug-13	18-Jun-14	Norfolk PD
12-Apr-13	02-Aug-13	DFS Central Lab
05-Dec-08	12-Apr-13	Franklin County SO
04-Aug-08	05-Dec-08	DFS Central Lab

INTOX EC/IR II
Quality Assurance Worksheet

IMS
mmB

Instrument Serial Number 010610 Worksheet Start Date 6/25/2015
Location Arlington Police Department
Address 1425 N. Courthouse Rd., Arlington, VA 22201
DFS Technician Heather Stanton License No. 18910
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 755 Reference Barometer (mm HG) 758
Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/16/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.301
0.300	0.298	0.289	0.306	Sample 2	0.300
Precision		sample min	sample max	Sample 3	0.299
0.002		0.299	0.301		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.099
0.100	0.100	0.097	0.103	Sample 2	0.099
Precision		sample min	sample max	Sample 3	0.099
0		0.099	0.099		

Dry gas standard Lot No. (with tank no.) AG509101-39

☐ Replaced dry gas standard (+O-ring)

☒ Installed at Location

☐ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

100
5 pads

Notes:

Instrument Serial Number

010610

Certification Date

06/26/15

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stark

Date

6/26/15

Issuing Analyst

Michelle Bullitt

Date

6/26/15SJA

AMS
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010610 Test Number: 1816

Test Date: 06/25/2015 Test Time: 11:30 EDT

Dry Gas Target: 0.298

Lot Number: AG506501-01 Exp Date: 03/06/2017

Tank Pressure: 517 psi Barometric Pressure: 756 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:31
CHK	0.301	11:32
BLK	0.000	11:33
CHK	0.300	11:34
BLK	0.000	11:35
CHK	0.299	11:36

Test Status: *Success*

Calibration CRC: 286C2E4B

LMS
mmb

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010610 Test Number: 1817

Test Date: 06/25/2015 Test Time: 11:40 EDT

Dry Gas Target: 0.100

Lot Number: AG509101-39 Exp Date: 04/01/2017

Tank Pressure: 1120 psi Barometric Pressure: 756 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:41
CHK	0.099	11:42
BLK	0.000	11:43
CHK	0.099	11:44
BLK	0.000	11:45
CHK	0.099	11:46

Test Status: *Success*

Calibration CRC: 286C2E4B




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

kmz
mmB

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 06/25/2015
TEST EQUIPMENT NUMBER 010610		

RESULTS: TIME SAMPLE TAKEN 11:56 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 25 Jun 2015 at 15:59

HMS
mmB

Test Results

Instrument Serial Number 010610

Test # 001818 Subject Test

Test Location 1 Department of
Test Date 25 Jun 2015

Test Location 2 Forensic Science
Test Time 11:50
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 25 Jun 2015 End Time 11:57

Result Time 11:56

Result Date 25 Jun 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:50

Data Type BLK

Sample Value 0.000

Sample Time 11:51

Data Type CHK

Sample Value 0.099

Sample Time 11:52

Data Type BLK

Sample Value 0.000

Sample Time 11:53

Data Type SUBJ

Sample Value 0.000

Sample Time 11:54

Data Type BLK

Sample Value 0.000

Sample Time 11:55

Data Type SUBJ

Sample Value 0.000

Sample Time 11:56

Data Type BLK

Sample Value 0.000

Sample Time 11:57

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG509101-39

Standard Expiration Date 04/01/2017

Tank Pressure 1120

Barometric Pressure 756 mmHg

Blow Sample Number 1 Blow Duration 3.32 sec

Blow Volume 1751 cc End-of-Blow Time 11:54

Blow Sample Number 2 Blow Duration 3.42 sec

Blow Volume 1766 cc End-of-Blow Time 11:56

Tamper Evident Stamp 7758ed00

Test Status Code 0

Test Status Success